



विदेह नगरपालिका नगर कार्यपालिकाको कार्यालय

पत्र संख्या :- २०८२१०८३

चलानी नं.:- १००३

विदेह नगरपालिकाको कार्यालय
तिनकौरिया चौक
धनुषा
स्वास्थ्य शाखा

मधेश प्रदेश, नेपाल

मिति : २०८३/०२/२२

विषय:- दरभाउ पत्र पेश गर्ने सम्बन्धि सूचना ।

श्री इच्छुक ईजाजत प्राप्त फर्म/सप्लायर्सहरु ।

प्रस्तुत विषयमा यस नगरपालिका अन्तर्गतको स्वास्थ्य संस्थाहरुको लागि यसै साथ संगलन सुचि अनुसारको औजार उपकरणहरु खरिद गर्नु पर्ने भएकोले यो सुचना प्रकाशन भएको मितिले ७ सात दिन भित्र प्रति एकाई दरभाउ खुलाई निम्न कागजातहरु संगलन राखि यस कार्यालयमा पेश गर्नु हुन अनुरोध छ ।

माँग बमोजिमको आवश्यक औजार उपकरणहरुको संख्या र स्पेसिफिकेशन (BOQ) यस पालिकाको वेबसाईटमा उपलब्ध छ ।

पेश गर्नु पर्ने कागजातहरु:-

- १.कम्पनी वा फर्म दर्ता प्रमाणपत्रको प्रतिलिपि म्याद ननाघेको ।
- २.व्यवसाय दर्ता र नविकरण प्रमाणपत्रको प्रतिलिपि ।
- ३.मु.अ.कर दर्ता प्रमाणपत्रको प्रतिलिपि ।
- ४.आ.ब. २०८१/०८२ को कर चुक्ता प्रमाणपत्रको प्रतिलिपि ।

नोट:- १.सम्पूर्ण कागजातहरु खामबन्दी गरी पेश गर्नु पर्नेछ ।

२.सुचि दर्ता नभएका फर्महरुले सुचि दर्ता गराई प्रस्ताव पेश गर्नु पर्नेछ ।

टोलराज उपाध्याय

उप सचिव अधिकृत
प्रमुख प्रशासकिय अधिकृत

बोधार्थ -

१.सूचना तथा प्रविधि शाखा:- नगरपालिकाको वेबसाईटमा अपलोड गर्नु हुन ।

हराम

विदेह नगरपालिका
नगरकायपालिकाको कार्यालय
तिनकीरिया धनुषा
आ.व. ०६/०६/०६३
विन्कीरिया, धनुषा

स्वास्थ्य संस्थाहरुको लागि उपकरण खरिद तथा वितरण कार्यक्रमको दररेट पेश गर्ने फारम (BOQ)

S.N.	Particular	Specification	Unit	Quantity	Rate	Total	Remarks
1	BP set	White Enamel 300 mmhg non-stop pin manometer, chrome-plated Brass Air Release valve.	Set	27			
2	Stethoscope	PVC tubing (Length 50-75 cm), a stainless steel or aluminium chestpiece with a tunable diaphragm and angled, soft-sealing eartips for acoustic	Set	27			
3	Autoclave	Capacity: 18 Liters, Chamber Material: Heavy-duty 304 Aluminium or stainless steel, Operating Temperature: 121 degree centigrade, Sterilization Time: adjustable, usually ranging from 0 to 60 minutes, Voltage and Power: 220v AC-240VAC, 50/60Hz, drawing 1500W to 2400W, water tanks: Typically includes a 3.5L to 4L clean Water tank and a corresponding wastewater tank, Accessories: includes Multiple tray holder.	Pcs	6			
4	ENT Set	chrome plated handle ty on/off switch, chrome plated otoscope head with removable 4* magnifying lens and inflator fitting, autoclavable polypropylene ear specula in 2.5mm, 3.5mm, 4.5mm, ophthalmoscope head with len wheel featurng 25 correctiv lens from -25 to +40dioptrs, TongueDepressor adopter, Hard plastic or raksin or leather box for ENT Diagnostic set	Pcs	9			
5	Dental Instruments Set	Included Mouth mirror, Dental probe, cotton pliers, periodontal probe, scalers, currettes, excavator, Dental forceps, Elevators, scissors	Set	6			


[Signature]

[Signature]

[Signature]

विदेह नगरपालिका
नगरकायपालिकाको कार्यालय
तिनकीरिया धनुषा

6	Footstep	Anti-slip platforms, sturdy steel or plastic frames and heights ranging from 12 to 30 inches	Pcs	9		
7	Weighing Machine Adult	capacity:180 kg, Mode of operation - manual, Display type- Analog, shape- Square	Pcs	18		
8	Weighing Machine Child	Pan type, 20-30kg maximum capacity, 5-10gram accuracy, removable cardle tray	Pcs	18		
9	Fetal Doppler Digital	Ultrasound frequency 2MHz, probe type: continuous Wave, FRH Measuring Range 50-240bpm	Pcs	9		
10	pulseoxymeter Fingertip	SpO2 Measurement Range-70 to 99%, SpO2 Accuracy: +-2% (at 80% to 99% saturation), pulse rate range: 30 to 240bpm, pulse rate accuracy: +-2 bpm	Pcs	15		
11	Nebulizer Machine	particle size: 1 to 5 microns, Nebulization Rate: 0.2ml to 0.4 ml/min., Medication cup capacity: 5 ml to 10 ml, Power source: standard 220v-240v AC adapters, weight: 200 gram up to 105 kg	Pcs	9		
12	Thermameter Digital	Measurement Range: 32 degree C. - 42.9 degree C.,	Pcs	102		
Total Amount						
Vat 13%						
Grand Total Amount						



Signature

prepared by

Signature

Signature

Signature
 Approved by
 Director, Health Services, BHM